

Paczków, 12 May 2015

Dear Sirs,

Chemical alcohol testers model ALCOTEST have been tested by Calibration Laboratory of Breath Analyzers Transcom International to the Customer's order. The testing was performed in accordance with standards NF X 20-702 June 2007 paragraphs 4.1÷ 4.4 and NF X 20-702 October 2014 paragraphs 6.1÷6.5.

The measurements have been performed for legal values specified in the highway code according to the standards NF X 20-702.

Sensitivity of chemical alcohol testers ALCOTEST:

No.	The mass concentration of alcohol in exhaled breath [%]	Sensitivity* [%]	Standards
1	0,2	96	NF X 20-702 June 2007
2	0,5	98,6	NF X 20-702 October 2014
3	0,8	100	NF X 20-702 October 2014

*sensitivity - the ratio of true positives to the sum of true positives and false negatives

The results presented refer only to the tested samples.

Yours sincerely,


Marta Królewska
Quality Manager